## **Technical Service Consultants Ltd**

Microbiology House, Fir Street, Heywood, Lancashire OL10 1NW, UK T: +44 (0)1706 620600 F: +44 (0)1706 620445

E:sales@tscswabs.co.uk W:www.tscswabs.co.uk

## **Product Specification Sheet**

**SS-L01**SwabSURE ListeriaP for direct detection of pathogenic Listeria by specific colour change

Product Code	SS-L01
Description	SwabSURE ListeriaP for direct detection of pathogenic Listeria by specific colour change
Intended Use	Direct detection of pathogenic Listeria
Components	Polypropylene tube, yellow polypropylene screw cap, Breakpoint foam tip swab, 150mm polypropylene tube, labels x2.
Contents	De-ionised Water, Chromogen for L. monocytogenes, Tween 80, Lecithin, Sodium Thiosulphate, Sodium Thioglycollate, Buffers to maintain pH - 7.2±0.2.
Method of Sterilisation	Sterilised by gamma irradiation at dosage of 25-53.4kGy
Shelf Life	12 months from date of manufacture
Storage Conditions	4-25°C. Store in dry and ventilated area; protect from light, heat and moisture in tightly closed original packaging
Precautions and Limitations	Restricted use for the presumptive detection of pathogenic Listeria from food contact and environmental surfaces. No other uses are authorised.
Pack Size	81 vials and swabs in a box
Other	Quality control testing details on final product: i) Validation of irradiation sterility. ii) Target Organism recovery & Final colour

This specification (and formulation where applicable) are intended for guidance only and may be adjusted to meet product defined criteria.

You should always refer to a professional with experience in your intended application if in any doubt.

Issue #: 7 Date: 01.05.2022 ProductSpec\_SS-L01