



Product Specification Sheet

TS/19-M250

Low ATP Swab with a polystyrene shaft and Dacron tip, in polypropylene outer tube

Product Code	TS/19-M250
Description	Low ATP Swab with a polystyrene shaft and Dacron tip, in polypropylene outer tube
Intended Use	Culture swab; general Laboratory use - also suitable for ATP measuring applications
Components	Polypropylene Tube, Blue HDPE cap, Polyester tip/polystyrene shaft swab, Non-toxic glue, Label.
Contents	Not applicable.
Method of Sterilisation	Sterilised by gamma irradiation at dosage of 25-53.4kGy
Shelf Life	24 months from date of manufacture
Storage Conditions	4-25°C. Store in dry and ventilated area; protect from light, heat and moisture in tightly closed original packaging
Precautions and Limitations	This swab will not fit in certain instruments. Please check with instrument manufacturer for compatibility
Pack Size	250 in a box
Other	Quality control testing details on final product: i) Validation of irradiation sterility ii) ATP compliance Test

This specification (and formulation where applicable) are intended for guidance only and may be adjusted to meet product defined criteria.

You should always refer to a professional with experience in your intended application if in any doubt.

Issue #: 5 Date: 24.04.2020

ProductSpec_TS19-M250